



6330 Hollister Avenue
Santa Barbara, CA 93117
Tel. (805) 681-9009
Fax (805) 681-0123
email: info@wyatt.com
www.wyatt.com

Certificate of Analysis

Sodium Chloride Solutions Kit

Sodium chloride (NaCl) in water: $dn/dc = 0.174$ mL/g at 658 nm; $dn/dc = 0.172$ mL/g at 690 nm

Wyatt p/n: 900080

Lot Number: 09032024

Expires: 01 Sep 2027

Storage: Room temperature

Standard analysis data:

Standard	CAS#	Purity	Concentration (mg/mL)
18 MΩ water	7732-18-5	100.00 %	NEAT
0.100 mg/mL NaCl Solution	7764-18-5	99.999 %	0.100
0.250 mg/mL NaCl Solution	7764-18-5	99.999 %	0.250
0.500 mg/mL NaCl Solution	7764-18-5	99.999 %	0.500
1.000 mg/mL NaCl Solution	7764-18-5	99.999 %	1.000
1.200 mg/mL NaCl Solution	7764-18-5	99.999 %	1.200
2.000 mg/mL NaCl Solution	7764-18-5	99.999 %	2.000
3.000 mg/mL NaCl Solution	7764-18-5	99.999 %	3.000
4.000 mg/mL NaCl Solution	7764-18-5	99.999 %	4.000
5.000 mg/mL NaCl Solution	7764-18-5	99.999 %	5.000

Certification: Michelle Chen

Michelle Chen
Senior Director, Research & Development

Date: 11 Oct 2024