





Certificate of Analysis

Batch DLS Size Standard (150 nm nominal radius)

Polystyrene latex spheres suspended in 10 mM sodium phosphate, pH 7.2, at 0.0001% w/v (~1×10⁸ particles/mL)

Wyatt p/n: 900174
 Lot Number: W150SD0425

Prepared by: Wyatt Technology, LLC
 Expires: 16 Jan 2027
 Storage: 2-8 °C after opening, DO NOT FREEZE



Standard characterization data by Wyatt DynaPro™ dynamic light scattering instruments:

Sample was measured neat at 25 °C solvent viscosity 0.90 cP.

R_h (Cumulants) = (141 ± 3) nm

R_h (Regularization) = (144 ± 5) nm

Certification: Sophia Kenrick
 Sophia Kenrick
 Director of Analytical Sciences

Date: 16 Jan 2026